


# 計量器静荷重検査 成績書

検査員 一般計量士 鬼丸忠士(登録番号 計測士 421-0218)  
 (〒421-0218) 静岡県焼津市下江留  
 株式会社 中央技術センター



社名	株式会社安芸砂利 安芸生コン工場										検査年月日										(〒421-0218) 静岡県焼津市下江留									
住所	三重県津市安濃町戸島2510番地										令和 6年 10月 8日										株式会社 中央技術センター									
名称	砂利1, 2			砂利3, 4			砂			セメント			水			混和剤 1			混和剤 2											
表示	SB2572-1 (4級)			SB2572-2 (4級)			SB2572-3 (4級)			SB2572-4 (4級)			SB2572-5 (4級)			SB2572-6 (4級)			SB2572-7 (4級)											
容量	2000-2kg			2000-2kg			2000-2kg			800-1kg			400-0.5kg			15-0.02kg			8-0.01kg											
感度	良			良			良			良			良			良			良											
往復	荷重	往	復	荷重	往	復	荷重	往	復	荷重	往	復	荷重	往	復	原器	往	復	原器	往	復	荷重	往	復						
計 量 器 の 読 み (kg)	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0									
	200	0	0	200	0	0	200	0	0	80	0	0	40	0	0	1	0	0	1	0	0									
	400	0	0	400	0	0	400	0	0	160	0	0	80	0	0	2	0	0	2	0	0									
	600	0	0	600	0	0	600	0	0	240	0	0	120	0	0	3	0	0	3	0	0									
	800	0	0	800	0	0	800	0	0	320	0	0	160	0	0	4	0	0	4	0	0									
	1000	0	0	1000	0	0	1000	0	0	400	0	0	200	0	0	5	0	0	5	0	0									
	1200	0	0	1200	0	0	1200	0	0	480	0	0	240	0	0	6	0	0	6	0	0									
	1400	0	0	1400	0	0	1400	-2	0	560	0	0	280	0	0	7	0	0	7	0	0									
	1600	-2	0	1600	0	0	1600	0	-2	640	0	0	320	0	0	8	0	0	8	-0.01	*									
	1800	-2	-2	1800	0	0	1800	-2	-2	720	0	0	360	0	0	9	0	0												
2000	-2	*	2000	0	*	2000	-2	*	800	-1	*	400	0	*	10	0	0													
																11	0	0												
																12	0	0												
																13	-0.02	0												
																14	0	-0.02												
																15	-0.02	*												
判定	合格 /			合格 /			合格 /			合格 /			合格 /			合格			合格 /											

合格判定基準：使用公差以内 ロードセル校正証明書番号：53-2320616 器物番号：No. YK5260004, YK5260005 有効期限：令和 7年 4月末日

適 合

